

Q-TRAK™ XP INDOOR AIR QUALITY MONITOR MODEL 7585



INTERFERING GASES AND CROSS SENSITIVITY

APPLICATION NOTE TSI-166 (US)

Interfering Gases and Cross Sensitivity

Cross sensitivity occurs when a sensor reacts to a gas that is not the target gas. This interfering gas causes a reaction in the sensor and displays a change to the readings even if the target gas is not present.

TSI® Incorporated gas sensors are built for optimum response to the target gas. Unfortunately, other “interfering” gases that may be present along with the target gas can create false signals.

Tables 1–8 show the average cross sensitivity for the most common interfering gases. The percent sensitivity is a measure of the magnitude of cross sensitivity; for example a 20% cross sensitivity means that 100 ppm of the interfering gas would read 20 ppm on the display (or $100 * 20\% = 20$ ppm).

Table 1: CO gas sensor cross-sensitivity to the most common interfering gases.

Carbon Monoxide (CO) Gas Sensor – 801401		
Interfering Gas	PPM applied	% Sensitivity
H ₂ S	5	<0.1
H ₂ @20c	100	<50
SO ₂	5	<0.1
NO ₂	5	<-2
NO	5	<-2
CL ₂	5	<0.1
C ₂ H ₄	100	<0.5
NH ₃	20	<0.1

Table 2: H₂S gas sensor cross-sensitivity to the most common interfering gases.

Hydrogen Sulfide (H ₂ S) Gas Sensor – 801402		
Interfering Gas	PPM applied	% Sensitivity
CO	5	<1
H ₂	100	<0.5
SO ₂	5	<15
NO ₂	5	<-20
NO	5	<3
CL ₂	5	<-8
C ₂ H ₄	100	<0.5
NH ₃	5	<0.1
CO ₂	5%	<0.1



Table 3: NO₂ gas sensor cross-sensitivity to the most common interfering gases.

Nitrogen Dioxide (NO₂) Gas Sensor – 801405		
Interfering Gas	PPM applied	% Sensitivity
CO	5	<-3
H ₂ S	5	<-80
H ₂	100	<0.1
SO ₂	5	<-3
NO	5	<5
CL ₂	5	<100
C ₂ H ₄	100	<1
NH ₃	20	<0.2
CO ₂	5%	<0.1

Table 4: CL₂ gas sensor cross-sensitivity to the most common interfering gases.

Chlorine (CL₂) Gas Sensor – 801400		
Interfering Gas	PPM applied	% Sensitivity
CO	400	<0.1
H ₂ S	20	<-300
H ₂	400	<0.1
SO ₂	20	<-8
NO ₂	10	100
NO	50	<3
C ₂ H ₄	400	<0.1

Table 5: NO gas sensor cross-sensitivity to the most common interfering gases.

Nitric Oxide (NO) Gas Sensor – 801404		
Interfering Gas	PPM applied	% Sensitivity
CO	5	<0.3
H ₂ S	5	<20 (after 3 minutes)
H ₂	100	<0.1
SO ₂	5	<4
NO ₂	5	<7 (after 3 minutes)
CL ₂	5	<4
NH ₃	5	<0.1
CO ₂	5%	<0.1
Halothane	100	<0.1

Table 6: O₃ gas sensor cross-sensitivity to the most common interfering gases.

Ozone (O₃) Gas Sensor – 801406		
Interfering Gas	PPM applied	% Sensitivity
CO	5	<-3
H ₂ S	5	<-80
H ₂	100	<0.1
SO ₂	5	<-3
NO	5	<5
CL ₂	5	<100
C ₂ H ₄	100	<0.1
NH ₃	20	<0.1
CO ₂	5%	<0.1
Halothane	100	<0.1

Table 7: CH₂O gas sensor cross-sensitivity to the most common interfering gases.

Formaldehyde (CH₂O) Gas Sensor – 801409		
Interference Gas	Concentration (ppm)	% Sensitivity
Carbon Monoxide	50	1
Ethyl Alcohol	2000	0.01
Acetic Acid	2000	-0.02
Ethylene	100	0.3
Methyl Alcohol	100	0.1
Isopropanol	100	0.1

Table 8: NH₃ gas sensor cross-sensitivity to the most common interfering gases.

Ammonia (NH₃) Gas Sensor – 801403		
Interference Gas	Concentration (ppm)	% Sensitivity
H ₂ S	25	88
Ethanol	200	0.5

Review of Tables 1–8, note that:

- Cross-sensitivity of interfering gases is temperature dependent. Some gas interferences will increase at higher temperatures while others will decrease.
- Cross-sensitivities were measured at 22°C for CO, H₂S, NO₂, CL₂, NO, and O₃ while CH₂O and NH₃ were measured at 20°C.
- Some interfering gases have a negative effect (for example NO₂ with H₂S sensors). This means that these interfering gases will decrease, rather than increase the signal. Beware that this could cause a situation where a target gas is at an alarm level but an interfering gas is preventing the detector from alarming.
- Cross-sensitivity of interfering gases vary with test environment.
- Cross-sensitivity of interfering gases may vary between batches of sensors.
- Some interfering gases do not simply reversibly interfere but actually poison the sensor, for example benzene or toluene will poison a H₂S sensor.
- Interfering gases act on the sensor in two ways:
 1. Most interfering gases (CO, H₂, H₂S) react on the working electrode, generating a current. The interfering signal rapidly stabilizes and is stable.
 2. Some interfering gases (C₂H₄, NO) modify the reference electrode potential, causing a potential shift on the working electrode. This variable interference stabilizes after typically 30 minutes.
- Chemical filters are used in the CO sensor. This chemical filter removes interfering gases by: adsorbing onto the chemical filter material, chemically absorbing the interfering gas, or catalytically reacting with the interfering gas.

These chemical filters have a finite lifetime that is different than the electrochemical gas cell lifetime. Therefore, CO sensor will show increased cross sensitivity to certain interfering gases after heavy exposure to these interfering gases.
- CO gas sensor has a H₂S filter with a capacity of 250,000 ppm-hrs.
- The NO₂ gas sensor has an O₃ filter with a capacity of >500 ppm-hrs.

References

Alphasense (www.alphasense.com) Application Note AAN 109-2



UNDERSTANDING, ACCELERATED

TSI Incorporated – Visit our website www.tsi.com for more information.

USA Tel: +1 800 680 1220
UK Tel: +44 149 4 459200
France Tel: +33 1 41 19 21 99
Germany Tel: +49 241 523030

India Tel: +91 80 67877200
China Tel: +86 10 8219 7688
Singapore Tel: +65 6595 6388



Distributed by:
Kenelec Scientific Pty Ltd
1300 73 22 33
sales@kenelec.com.au
www.kenelec.com.au