

RECOMMENDATIONS FOR USE OF A 40 µm FUSED SILICA CAPILLARY WITH MODEL 3480 ELECTROSPRAY AEROSOL GENERATOR

TECHNICAL NOTE

Elisabeth M. Loecken, Ph.D.

*TSI Incorporated, 500 Cardigan Road, Shoreview, MN 55126 USA
(29 December 2011)*

Introduction

TSI's Model 3480 Electro Spray Aerosol Generator produces monodisperse aerosol with a nominal droplet size of 150 nm using a fused silica capillary with an inner diameter of 25 µm. Depending on the specific analyte and buffer solution, these capillaries can readily clog, complicating sample analysis. Generally, the larger the particle to be electro sprayed, the more likely a clog will form, although surface chemistry also plays a role in the clogging process. Many users have found success electro spraying analytes using a capillary with an inner diameter of 40 µm. TSI has yet to fully characterize the 40 µm fused silica capillary (P/N 3900126), but general instrument settings have been determined in order to produce a stable cone spray.

Recommended Instrument Settings

Instrument Setting	25 µm Capillary	40 µm Capillary
Differential Pressure	3.7 psid	2.7 psid
Voltage	2.0 kV	2.0 kV
Current	(-)280-320 nA	(-)340-400 nA*
Air Flow	1.0 L/min	1.4 L/min
CO ₂ Flow	0.1 L/min	0.1 L/min

Notes:

1. The 40 µm capillary will result in higher particle concentrations due to the larger volume of liquid being electro sprayed.
2. The nominal droplet size produced by a 40 µm capillary has not been determined.

*The electrical current depends on buffer constitution. The recommended values have been observed for 20 mM ammonium acetate buffer (pH 8.0).





UNDERSTANDING, ACCELERATED

TSI Incorporated – Visit our website www.tsi.com for more information.

USA Tel: +1 800 874 2811
UK Tel: +44 149 4 459200
France Tel: +33 4 91 11 87 64
Germany Tel: +49 241 523030

India Tel: +91 80 67877200
China Tel: +86 10 8251 6588
Singapore Tel: +65 6595 6388



Distributed by:
Kenelec Scientific Pty Ltd
1300 73 22 33
sales@kenelec.com.au
www.kenelec.com.au