



Ammonia (NH_3) Module

Model 7591-11

Operation Instructions

1. **Power On:** Hold down power button on left side of module for 1 second.
2. **Power Off:** Hold down power button on left side of module until you see a light sequence of green, yellow, red. That indicates module is off.
3. **Charging:** Plug USB-C into bottom of module and into power source. A blinking green light indicates module is charging.
4. **Connecting to a Smart Station:** Power on module by holding down the circle on the left for 1 second. A blue LED light will start to blink indicating Bluetooth® is on.
 - The module will appear in the list on the Smart Station.
If not, hit Refresh.
 - Press desired module on screen to connect to Smart Station.
A green check mark indicates a successful connection.
 - Press Done

Now you're ready to take measurements.

The OmniTrak™ Smart Station can be used in conjunction with any OmniTrak™ Module to provide immediate feedback as to the conditions in the immediate area. Take measurements, create reports, and analyze data provided to help improve conditions. This instrument is a handheld device, not to be worn on the body, or near an individual's head.

Dispose of battery in accordance with local rules and regulations.

What's Included in the Box

- Device
- USB-C Cable

NOTE: All modules, by default, are shipped in 'Sleep Mode' to protect the battery. It will not be possible to power on the Module, using the side button, until 'Sleep Mode' is disabled. To disable, plug the module into a power source and charge the module for a few minutes. This process is required only once.

Ammonia (NH₃) Module

Model 7591-11

Specifications

Ammonia Sensor Specifications

Sensor Type	Electrochemical
Range	0 - 100 ppm
Accuracy	+/- 10 ppm
Resolution	0.1 ppm
*Response Time	300 seconds

Measurement specifications apply at ambient conditions of 21 +/- 5 °C temperature, 98.6 +/- 5 kPa pressure, and 50 +/- 10% relative humidity.

*(Typical) time to 90% of final value

AC and DC Power Requirements

AC/DC Power Supply	100 to 240 VAC, 50/60 Hz, 0.3A
Device Input Power	5 VDC, 2A, 10W

Environmental/Installation Requirements

Maximum Altitude	3000 meters (10,000 feet)
Pollution Degree	2
Installation Category	I
Operating Temperature	0-30 °C dew point
Storage Temperature	-20-35 °C dew point
Humidity	0% - 95% (non-condensing)

Compliance

FCC ID: 2A22JOTMODULES IC ID: 28101-OTMODULES

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

1. This device may not cause harmful interference.
2. This device must accept any interference received, including interference that may cause undesired operation.

Per FCC 15.21 Changes or modifications not expressly approved by TSI will void the product warranty and may void the user's authority to operate the equipment.

Compliance

This device contains license-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's license-exempt RSS(s). Operation is subject to the following two conditions:

1. This device may not cause interference.
2. This device must accept any interference, including interference that may cause undesired operation of the device.

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes:

1. l'appareil ne doit pas produire de brouillage, et
2. l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

To satisfy RF exposure requirements, this device and its antenna may be operated near a person's body and the extremities, but must be operated with a separation distance of at least 20 cm from all person's head.

Pour satisfaire aux exigences d'exposition RF, cet appareil et son antenne peuvent être utilisés à proximité du corps et des extrémités d'une personne, mais doivent être utilisés avec une distance de séparation d'au moins 20 cm de la tête de toute personne.

Ammonia (NH₃) Module

Model 7591-11



WARNING

- **DO NOT** attempt to replace the battery in this instrument.
- **DO NOT** abuse the battery in any way as the battery may rupture or catch fire.
- **DO NOT** use a substitute or non-rechargeable battery in this instrument.
- **DO NOT** short-circuit, incinerate, dismantle or mutilate Lithium ion batteries
- **DO NOT** expose to water or heat.
- **DO NOT** use any battery which shows signs of damage, such as bulging, swelling, a swollen plastic wrap, liquid in the plastic wrap, etc.



Knowledge Beyond Measure.

TSI Incorporated - Visit our website www.tsi.com for more information.

USA

Tel: +1 800 874 2811

UK

Tel: +44 149 4 459200

France

Tel: +33 1 41 19 21 99

Germany

Tel: +49 241 523030

India

Tel: +91 80 67877200

China

Tel: +86 10 8219 7688

Singapore

Tel: +65 6595 6388

P/N 5003101 Rev C

©2024 TSI Incorporated

Printed in U.S.A.



Distributed by:

Kenelec Scientific Pty Ltd
1300 73 22 33
sales@kenelec.com.au
www.kenelec.com.au

Specifications are subject to change without notice.

Wi-Fi is a registered trademark of Wi-Fi Alliance®.

The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by [licensee name] is under license. Other trademarks and trade names are those of their respective owners.

TSI, and the TSI logo are registered trademarks of TSI Incorporated in the United States and may be protected under other country's trademark registrations.

6176981034